



UKCA Declaration of Conformity

We H2flow Controls, Inc.
of 3545 Silica Road
Unit F
Sylvania, OH 43560
USA

in accordance with the following Directives:

- The Electrical Equipment (Safety) Regulations 2016
- The Electromagnetic Compatibility Regulations 2016
- The Restriction of the Use of Certain Hazardous Substances in Electrical & Electronic Equipment (Amendment) Regulations 2021

hereby declare under our sole responsibility that:


Equipment PD100 Power Monitor
Model number(s) D304-001

are in conformity with the applicable requirements of the following documents:

<u>Reference</u>	<u>Title</u>	<u>Edition (Date)</u>
IEC 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use	3.0 (2010-06)
IEC 61326-1	EMC requirements for equipment designed for measurement, control and laboratory use	2.0 (2012-07)
IEC 61000-4-2	Electrostatic discharge (ESD) immunity test	2.0 (2008-12)
IEC 61000-4-3	Radiated, radio-frequency, electromagnetic field immunity test	3.2 (2010-04)
IEC 61000-4-4	Electrical fast transient/burst immunity test	3.0 (2012-04)
IEC 61000-4-5	Surge immunity test	3.0 (2014-05)
IEC 61000-4-6	Immunity to conducted disturbances, induced by radio-frequency fields	4.0 (2013-10)
IEC 61000-4-8	Power frequency magnetic field immunity tests	2.0 (2009-09)
IEC 61000-4-11	Immunity to voltage dips, interruptions and variations.	2.0 (2004-03)
CISPR-11	Conducted and Radiated Emissions	6.0 (2016-06)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all applicable Essential Requirements of the Directives.

Signed by:


Name: Gavin Hackett
Position: President
Done at: Sylvania, OH USA
On: 08/10/2022



Document number: 5743

H2flow Controls, Inc.

3545 Silica Road
Unit F
Sylvania, OH 43560
USA

tel +1 888 635 0296
fax +1 419 517 9900
email sales@h2flow.net
web www.h2flow.net

Registration number: 7115948

Progress Through Innovation™